Strassburger Filter GmbH + Co. KG

CERTIFICATE OF COMPLIANCE

(according to DIN 50049 - 2.3 / EN 10204 - 2.3)

SF - SS F 400 x 400 mm

Batch No. 233604606

Production date: 19-07-18 (yy.mm.dd)

We accept sole responsibility in stating that the product conforms to the presently valid and agreed specifications.

Depth filter sheets are products for solid/liquid separation in sterilizing or microorganism-reducing filtration, polishing or coarse filtration.

		Average batch value
Water flow rate ∆p 100 kPa	L/m²/mín	108
Weight per unit area	g/m²	1253
Thickness	mm	3,8
Ash content	(%)	46 Approx. values
Extractable ions Soluble in: acetic acid (5%) e.g.	(mgm ⁻²)/(mgkg ⁻¹) Ca Fe Al	1500 15 100

The tests are carried out as per the methods of the European Association of Depth Filtration e.V. or in accordance with company-specific test procedures.

Heavy metals as per recommendation XXXVI/1 within the Lebensmittel- und Futtermittelgeetzbuch - LFGB - (Foodstuffs and Animal Feed Code (ppm)

< 50

Bacteria retention capability with Serratia marcescens: LRV (Log Reduction Value)

(according to the method of the European Association of Depth Filtration e,V.)

6

The product conforms to the recommendation XXXVI/1 and meets the requirements of the Lebensmittel- und Futtermittelgesetzbuch - LFGB- (German Foodstuffs and Animal Feed Code) especially §§ 5, 30 and 31. It may without any reservation be used for cold filtration of foodstuffs.

Main components:

cellulose

Components < 3 %:

resins corresponding to the recommendation XXXVI resp.

natural diatomaceous earth periite

21 CFR polyolefine fibres

Safe filtration requires that all process parameters relevant to a special application have been validated. The above statements do not rescind from the obligation to perform an incoming goods control. All above information are based on our present state of knowledge and do not claim to be complete. The use resp. handling of our products is in the buyer's responsibility also with regard to possible protective rights of third parties.

The depth filter sheets should be used within 3 years of production

Westhofen, 23.08.2019

This document has been made by computer and is valid without a signature.

233604606.docx